



## EUROPEAN MEDICAL DEVICE REGULATION

### Declaration of Conformity

As Legal Manufacturer, we

3M Deutschland GmbH  
Health Care Business  
Single Registration Number: DE-MF-000012859  
Carl-Schurz-Str. 1  
41453 Neuss  
Germany

hereby declare under our sole responsibility that the following CE marked device(s)

Trade Name	Express XT
Intended Purpose	Dental impression material
Reference	36891, 36892, 36893, 36894, 36895, 36896, 36974, 36975, 36976, 36977
Basic UDI-DI	06082232761020000000004DJ

is/are classified per rules 4 and 19 of Annex VIII of the Medical Device Regulation (EU) 2017/745, as Class IIa devices in accordance with *Annex IX and all other* applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

We hereby declare under our sole responsibility that the following CE marked accessory(ies) which are intended to be used together with the above medical device(s)

Accessory(ies)	Reference	Basic UDI-DI	Rule(s) of Annex VIII	Class
Pentamix 3	77871, 77872, 77873, 77874, 77875, 77876, 77878	06082232761020000000095EF	13	I
Including its component: Penta Universal Cartridge for Pentamix 3	71545	06082232761020000000095EF	13	I



Pentamix Lite	77901, 77902, 77903, 77904, 77905, 77906, 77907, 77908, 77909, 78000	0608223276102 0000000095EF	13	I
Including its component: Penta Universal Cartridge for Pentamix Lite	77944	0608223276102 0000000095EF	13	I
Penta Mixing Tips – Red	77919, 77949	0608223276102 0000000094ED	1	I
Garant Dispenser 1:1/2:1	77580	0608223276102 0000000015DP	1	I
Garant Mixing Tip – Yellow	71452	0608223276102 0000000016DR	5	I
Garant Intraoral Tip - Yellow	71462	0608223276102 0000000017DT	5	I
Intra-oral Syringe – Green	71482, 71505, 71506	0608223276102 0000000009DU	5	I
Express XT Intra-oral Syringe	71483	06082232761020 0000000009DU	5	I
VPS Tray Adhesive	7307	0608223276102 0000000003DG	5	I

is/are classified according to Annex VIII of the Medical Device Regulation (EU) 2017/745 in accordance with all applicable provisions of the REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL concerning medical devices.

This declaration is made based on the quality assurance certificate  
EU QMS Certificate (MDR): G10 078535 0040  
Issued by: TÜV SÜD Product Service GmbH (identification no. 0123)

3M Deutschland GmbH self-declares conformity of Pentamix 3 and Pentamix Lite with Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment and compliance to the requirements of EN IEC 63000:2018.

Dr. Desi W. Soegiarto  
Manager Regulatory Medical Devices  
3M Deutschland GmbH

Seefeld, April 7, 2022  
Location/Date